



Agilent GPC-Addon Rev. A.02.02

Sample : b2s
Injection Date : 13-Aug-18, 09:51:34
Calibration File : G:\GOZARESHAT\CAL\CAL-9704-01.CAL
Calibration Date : Tuesday 18/07/03 09:40:12
Baseline from : 15.849 min
Integration from: 15.849 min
MHK - A (Cal.): 0.000000E+0
Eluent : THF
Concentration : 1.000 g/l
Column 1 : PLgel OH-MIXED C
Detector 1 : RID A, Refractive Index Signal
Operator : IPPI

Baseline to : 18.886 min
Integration to : 18.886 min
MHK - K (Cal.): 1.000000E+0 ml/g
Flowrate : 1.000 ml/min
Inject volume : 20.000 ul
Temperature : 29.990 C
Delay volume : 0.000 ml
Acquisition interval : 0.430 sec

rid1A

Mn : 1.6321e3 g/mol
Mw : 1.8762e3 g/mol
Mz : 2.1876e3 g/mol
Mv : 0.000000 g/mol
D : 1.1496e0
[n]: 0.000000 ml/g
Vp : 1.7503e1 ml
Mp : 1.6366e3 g/mol
A : 3.8940e3 ml*V
< 397 0.00
w% : 100.00
> 8895 0.00